Calibration facilities and sensor development:

There are two set-ups available: (A) with fluoroinert bath; (B) Dry process calibrator



Instrumentation:

- Platinum Resistance Thermometer TINSLEY [25Ω; accuracy ± 0.005 K]
- ASL Metal Block Calibrator B140
- Digital Multimeter Keithley, 195 A
- Multimeter Keithley, Model 2000 with scanner card



DAQ system used to monitor data from

- Reference thermometer TINSLEY
- Individual NTC sensors

Stability of the ASL calibrator is better then ± 0.03 K



- Special temperature probes have been be developed and used measurements at CERN
- Studies of the side effects on sensors are conducted (heat sinks for leads, influence of convection and radiation, etc.)

• Evaluation of irradiation effect on the Pt 100 sensor behavior have been determined





• Irradiation doses used:

9.7E12 +/-12 1.3E14 +/-9% and 9.7E14 +/-7% [p/cm2] (irradiation did not significantly influence the accuracy of the sensors; similar conclusion announced later by Wuppertal concerning the NTC sensors)

 The graph displays the absolute deviations of the noncalibrated sensor temperature readings from the correct values of temperature measured by the reference PRT TINSLEY within the investigated temperature range from -55 °C to +30°C.

- Calibration facilities and procedures were verified during several calibration measurements with Pt100, Pt1000 and NTC sensors
- Facilities for pressure sensor calibration are also available
 - Precise micro-manometer (0-600 mm H₂O),
 - Rosemount Pressure Transmeter E1151 (0-2500 mm H₂O),
 - Pressure transducer Druck PDCW 32 (0-3.5 bar)]
- Facilities for flow-meters calibration /liquid phase/ available